



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-244.10.08



Product Service

EC Certificate

Production Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex V

(Devices in class I in sterile conditions, sterilised systems or procedure packs)

No. G2S 012974 0609 Rev. 00

Manufacturer

B. Braun Melsungen AG

Carl-Braun-Str. 1
34212 Melsungen
GERMANY

Product Category(ies):

**Accessories for Angiography, Surgery,
Angiography / Atherectomy and Haemodynamic
Monitoring
(class I sterile)
Procedure Kits, Monitoring sets for invasive
physiological pressure measurement
(article 12 systems and procedure packs)**

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for manufacture in accordance with MDD Annex V. This quality assurance system covers those aspects of manufacture concerned with securing and maintaining sterile conditions of the respective devices / device categories and conforms to the requirements of this Directive. It is subject to periodical surveillance. See also notes overleaf.

Report No.:

713160243+713160242

Valid from:

2019-10-08

Valid until:

2024-05-26

Date,

2019-10-08

Stefan Preiß

Head of Certification/Notified Body



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-244.10.08



Product Service

EC Certificate

Production Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex V

(Devices in class I in sterile conditions, sterilised systems or procedure packs)

No. G2S 012974 0609 Rev. 00

Facility(ies):

B. Braun Melsungen AG Vascular Systems
Sieversufer 8, 12359 Berlin, GERMANY

B. Braun Melsungen AG Vascular Systems
Mistelweg 2, 12357 Berlin, GERMANY

AESCLAP CHIFA Sp. z o.o.
ul. Tysiaclecia 14, 64-300 Nowy Tomysl, POLAND

./.