

KARTA KONTROLI JAKOŚCI



produkt: Selenite F Broth

Numer katalogowy:	Numer serii:	20241007	Rodzaj podłoża:	wybiórcze
6030TB5	Data ważności:	02.04.2025	UDI-DI (GTIN)	5904708003217

Skład podłoża w g/l:

Trypton	5,0 g
Laktoza	4,0 g
Selenin sodu	4,0 g
Fosforan sodu	10,0 g

Suplementy g/l:

Ogólna charakterystyka podłoża:

	oczekiwany wynik:	uzyskany wynik:
kolor	jasnośłomkowy*	zgodny
klarowność / jednorodność	klarowny	zgodny
konsystencja	płynne	zgodny
pH	7,0 ± 0,2	7,0

Kontrola jałowości:

	oczekiwany wynik:	uzyskany wynik:
czas inkubacji: 3 - 5 dni	brak wzrostu	zgodny
temp. inkubacji: 37 i 25 ± 2°C		

Kontrola wzrostu:

czas inkubacji:	18-24h
temp inkubacji:	35 ± 2°C

Mikroorganizm:

Kryteria:

Wynik:

<i>Salmonella typhimurium</i> ATCC 14028	zmętnienie (2), ceglasta barwa podłoża	zgodny
<i>Escherichia coli</i> ATCC 25922	brak wzrostu (0)	zgodny

Kontrola jakości przeprowadzona zgodnie z zaleceniami ISO 11133:2014.

*** Mogą pojawić się ceglasteroczerwone percipitaty, które nie wpływają na jakość odczytu. Percipitaty znikają po wstrząśnięciu probówki.**

Data wystawienia certyfikatu: 11.10.2024

Badania wykonał: Laboratorium wewnętrzne

Dokument wystawił:

Krąg 4A; 83-200 Starogard Gdański
www.grasobiotech.pl
tel. + 48 (58) 562 30 21



Production Department
Leśna 1, Owidz
83-211 Jabłowo



P8F2 wer. 2.0 z 27.09.2017

CERTIFICATE OF ANALYSIS



product: Selenite F Broth

Article:	Lot No.:	20241007	Medium type:	selective
6030TB5	Expiry date:	02.04.2025	UDI-DI (GTIN)	5904708003217

Formula/Liter:

Tryptone	5,0 g
Lactose	4,0 g
Sodium selenite	4,0 g
Sodium phosphate	10,0 g

Supplements/Liter:

Characteristics of medium:

colour
clarity / homogeneity
consistency
pH

Spec. Values:

light straw
clear
liquid
7,0 ± 0,2

Batch Values:

conforms
conforms
conforms
7,0

Sterility control:

time: 3 - 5 days
temperature: 37 and 25 ± 2 °C

Spec. Values:

no growth

Batch Values:

conforms

Incubation temperature and time:

time: 18-24h
temperature: 35 ± 2 °C

Microorganism:

Criteria:

Results:

<i>Salmonella typhimurium</i> ATCC 14028	turbidity (2), orange medium	conforms
<i>Escherichia coli</i> ATCC 25922	no growth	conforms

Quality control was executed according ISO 11133:2014.

Date of revision: 11.10.2024

Test performed by: Internal Laboratory

Krąg 4A; 83-200 Starogard Gdański
www.grasobiotech.pl
tel. + 48 (58) 562 30 21



Production Department
Leśna 1, Owidz
83-211 Jabłowo



P8F2 wer. 2.0 z 27.09.2017