



# Certificate of Analysis



**Catalog No.:** 257079  
**Product Description:** Group B Streptococcus Differential Agar (Granada Medium)  
**Lot Number:** 4282360  
**Expiry Date:** 31. December 2024  
**Date of Manufacture:** 15. October 2024  
**Date of Release:** 23. October 2024  
**Storage Temperature:** 2 - 8° C

## Physical Characteristics

Plate type	90mm
Media colour / appearance	Light greyish, opaque
pH	7.4 ± 0.2
Sterility (7 days at 28°C)	Within set limits

## Microbiology

Organism	Strain No.	Expected growth (CFU/ml or growth intensity)	Actual growth (CFU/ml growth intensity or conformity)	Colony size and appearance
Streptococcus agalactiae	ATCC 12386	1E3 - 1E5	9.5E3	Small to medium-sized; orange
Streptococcus agalactiae	Wild 27	1E3 - 1E5	9.6E3	Small to medium-sized; pale orange
Streptococcus pyogenes	ATCC 19615	Complete to partial inhibition	Partial inhibition	Very small; colourless to grey
Enterococcus faecalis	ATCC 29212	1E3 - 1E5	1.8E3	Medium-sized; colourless to grey
Staphylococcus aureus	ATCC 25923	Complete inhibition	Conforms	N.A.
Proteus mirabilis	ATCC 43071	Complete to partial inhibition	Complete inhibition	Swarming completely inhibited; colourless
Bacteroides fragilis	ATCC 25285	Inhibited to trace	Complete inhibition	N.A.

Incubation: 18 - 24h, 35 - 37°C; anaerobic

We confirm that the above-mentioned product lot meets the specification.

**Manufactured** by Becton Dickinson GmbH · Tullastrasse 8 – 12, D-69126 Heidelberg (Germany)  
Phone: +49 62 21 30 50 · Fax: +49 62 21 30 52 16  
E-mail: [Reception\\_Germany@europe.bd.com](mailto:Reception_Germany@europe.bd.com)

Dr. Diana Popovici  
Quality Control Team-Leader

F-57079.07