



Certificate of Analysis



Catalog No.: 254060 / 254089
Product Description: BD Chocolate Agar (GC II Agar with BD IsoVitaleX™)
Lot Number: 4254471
Expiry Date: 16. December 2024
Date of Manufacture: 23. September 2024
Date of Release: 30. September 2024
Storage Temperature: 2 - 8° C

Physical Characteristics

Plate type	90mm
Media colour / appearance	Brown, opaque
pH	7.3 ± 0.2
Sterility (7 days at 28°C)	Within set limits

Microbiology

Organism	Strain No.	Expected growth (CFU/ml or growth intensity)	Actual growth (CFU/ml or growth intensity)
Neisseria gonorrhoeae	ATCC 43069	1E3 - 1E5	1.8E4
Neisseria meningitidis	ATCC 13090	1E3 - 1E5	8.9E3
Haemophilus influenzae	ATCC 10211	1E3 - 1E5	4.4E3
Staphylococcus epidermidis	ATCC 12228	1E3 - 1E5	5.5E3
Streptococcus pneumoniae	ATCC 6305	1E3 - 1E5	3.3E3
Granulicatella adiacens	DSM 9848	1E3 - 1E5	3.2E4

Incubation: 24-48h, 35-37°C, CO₂ - atmosphere

We confirm that the above mentioned product lot meets the specification.

Manufactured by Becton Dickinson GmbH · Tullastrasse 8 – 12, D-69126 Heidelberg (Germany)
Phone: +49 62 21 30 50 · Fax: +49 62 21 30 52 16
E-mail: Reception_Germany@europe.bd.com

Dr. Diana Popovici
Quality Control Team-Leader