



## • Certificate of Analysis

Product Name : **MICRONAUT-S Anaerobes MIC**

Part No.: E1-085-040  
LOT No.: 240215085A  
Use by: 2026-02-15  
Date of Production: 15.02.2024

Das oben genannte Produkt wurde mit den folgenden Stämmen in unserem Qualitätskontroll-Labor nach dokumentierten Kontrollverfahren geprüft.  
Die Ergebnisse entsprechen den spezifizierten Kontrollparametern.

*The above product has been verified with the following strains in our quality control laboratory according to documented control procedures.  
The results correspond to the specified control parameters.*

*List of strains:*

*Staphylococcus aureus ATCC 29213  
Escherichia coli ATCC 25922  
Pseudomonas aeruginosa ATCC 27853  
Enterococcus faecalis ATCC 29212  
Bacteroides fragilis ATCC 25285  
Bacteroides thetaiotaomicron ATCC 29741*

Die eingesetzten Kontrollverfahren sind festgeschriebene Anweisungen unseres internen Qualitätsmanagementsystems gemäß den Anforderungen der EN ISO 13485.

*The control procedures used are documented methods of our internal quality management system according to the requirements of EN ISO 13485.*

### Remarks:

QC release date: 19.02.2024

Bruker Daltonics GmbH & Co. KG, Germany  
Fahrenheitstr. 4 – 28359 Bremen  
Mail: micronaut.support@bruker.com  
phone +49-421-2205-0

Julia Bodemann  
Head of Quality Control

Sonja Lehmann  
Quality Management Representative