

TIPO DOCUMENTO

CONTROLLO SOPRAVVIVENZA BATTERICA SPOT



TITOLO

REV 00
DEL 02.2007
Page 1 of 1

COD. PRODUCT	LOT	PRODUCT DESCRIPTION	Exp. Date	Preparation date
111C.USE	AHT400	Transystem Stuart w/o charcoal	11/2009	04/2008

Testing temperature

22°-30°C

	T=0	T=24h			T=96h		
STRAIN A	++++	++++	++++	++++	++++	++++	++++
STRAIN B	++++	++++	++++	++++	++++	++++	++++

	KEY
	++++: confluent colonies /+++ : semi-confluent colonies /++ : 2000-4000 colonies / + : 200-2000
	ACCEPTABILITY LIMITS
	Strain A: <i>Shigella flexneri</i> ATCC 12022 Strain B: <i>Salmonella typhimurium</i> ATCC 14028 T=0 ≥ +++ T=24h ≥ 50 CFU/plate

Chief of Microbiology lab
Paroni Roberto