



Procedure Cancellation Notice

Concerns: open procedure No F2/40/2025/ZP for delivery of a device for plasma cleaning of substrates.

Acting pursuant to Art. 260 of the Public Procurement Law Act (Journal of Laws of 2019, item 2019 with subsequent amendments), hereinafter referred to as "PPL", the Contracting Authority hereby notifies about cancellation of the procedure.

Legal grounds: Art. 255 Par. 2) of the PPL, all requests to participate or tenders submitted are subject to rejection

Factual grounds: in the procedure, the bid was submitted by the Contractor: PIE Scientific LLC, 3209 Whipple Rd., Union City, CA 94587, USA.

The bid was rejected on the basis of Article 226 par. 1 point 5 - its content is inconsistent with the specification of essential terms of the reference.

The bid does not comply with the Contracting Authority's requirements specified in Appendix No. 1 to the Terms of Reference (ToR) in the following points:

- 13.4 - Equipped with Faraday cage – The Contractor did not confirm but stated „ *No, Since Tergeo-Plus has true downstream mode with remote source, there is no need to have faraday cage because the samples are not exposed to plasma or rf field in the downstream mode.*”
- 16.7 - Oxygen pressure regulator for compressed gas bottles. With shut off valve and ¼" compressed tubing connector. 0-50PSI output - The Contractor did not confirm but stated: „ *No, please purchase it locally within Poland for compliance*”

Przedmiot zamówienia będzie realizowany w ramach projektu: Inwestycja: A2.4.1 Inwestycje w rozbudowę potencjału badawczego Krajowego Planu Odbudowy i Zwiększania Odporności, Przedsięwzięcie: Centrum Kompetencji Mikroelektronika i Fotonika, Nr umowy KPOD.01.18-IW.03-0011/23



Rzeczpospolita
Polska

Sfinansowane przez
Unię Europejską
NextGenerationEU



Strona 1 z 2

- 16.8 - Argon pressure regulator for compressed gas bottles. With shut off valve and ¼" compressed tubing connector. 0-50PSI output - The Contractor did not confirm but stated: „ *No, please purchase it locally within Poland for compliance*”
- 16.9 - Nitrogen pressure regulator for compressed gas bottles. With shut off valve and ¼" compressed tubing connector. 0-50PSI output - The Contractor did not confirm but stated: „ *No, please purchase it locally within Poland for compliance*”.

